

Page: 1

Date: 12/29/23 at 11:32 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	OXY4121	?		
NITRIC ACID 0.5% V/V SOLUTION				
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	12/29/23			
Tester	AROS			
Recorded	12/29/23			
V/V% NITRIC ACID		0.50	0.49 - 0.51	0.50
		V/V%		
Sample ID				
Tested	12/29/23			
Tester	12/29/23 AROS			
	12/29/23			
Recorded	12/29/23	1.00079	1.00068 - 1.00091	1,00000
Actual Density		1.00078	1.00066 - 1.00091	1.00090
		GM/ML		
Sample ID				
Tested	12/29/23			
Tester	AROS			
Recorded	12/29/23			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	12/29/23			
Tester	AROS			
Recorded	12/29/23			
INSTRUMENTS USED D	URING	INSTRUMENT		GC,VF-A14
PREPARATION				•
Sample ID				
Tested	12/29/23			
Tester	AROS			
Recorded	12/29/23			
EXPIRATION DATE	12/23/23	MM/DD/YY		01/30/25
EXPIRATION DATE		MM/DD/YYYY		01/30/23
Sample ID				
Tested	12/29/23			
Tester	AROS			
Recorded	12/29/23			
Lot # 31229				
Made 12/29				
INIAGE ==7=3				

Page: 2

Date: 12/29/23 at 11:32 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager