

Page: 1

Date: 12/29/23 at 2:49 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	AME 0 0 4 1	F.		
_	IPA/TO	LUENE 75:25 V/V SOLUTION		
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES	<u> </u>	YES
FILE		YES		
Sample ID				
Tested	12/29/23			
Tester	JRIO			
Recorded	12/29/23			
TURNED YELLOW DURING		PASS		PASS
VERIFICATION		PASS		
Sample ID				
Tested	12/29/23			
Tester	JRIO			
Recorded	12/29/23			
Actual Density		0.80571	0.80448 - 0.80695	0.80595
		GM/ML		
Sample ID				
Tested	12/29/23			
Tester	JRIO			
Recorded	12/29/23			
APPEARANCE AND COL	.OR	CLEAR WATER WHITE		PASS
		APPEARANCE		
Sample ID				
Tested	12/29/23			
Tester	JRIO			
Recorded	12/29/23			
INSTRUMENTS USED DURING		LAB EQUIPMENT USED		M55,T59,VF12-87
PREPARATION		INSTRUMENT		
Sample ID				
Tested	12/29/23			
Tester	JRIO			
Recorded	12/29/23			
EXPIRATION DATE		MM/DD/YYYY		01/30/2025
		MM/DD/YYYY		
Sample ID				
-	12/29/23			
Tester	JRIO			
Recorded	12/29/23			
REFRACTIVE INDEX		1.4020	1.4000 - 1.4040	1.4006
		RI		
Sample ID				
Tested	12/29/23			
Tester	JRIO			
	12/29/23			
Lot # 312282				
Made 12/29,	/ 43			



Page: 2

Date: 12/29/23 at 2:49 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager