

Page: 1

Date: 12/27/23 at 2:02 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

		Custo	mer PO #:	
Packaged Produc	t: 205-12F	1		
	VERSENA	TE HARDNESS TITRATING S	SOLUTION 0.01 MOLAR	
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES	<del>.</del>	YES
FILE		YES		
Sample II	)			
Tested	12/27/23			
Tester	MHUS			
Recorded	12/27/23			
ACTUAL MOLARITY		0.0100	0.0095 - 0.0105	0.0098
		MOLARITY (M/L)		
Sample II	)			
Tested	12/27/23			
Tester	MHUS			
Recorded				
Prepared and Certified to API 13A		YES		YES
		YES		
Sample II	)			
Tested	12/27/23			
Tester	MHUS			
Recorded				
Calibration Correction		-12 TO + 12 MG/KG		YES
mg/KGA		MG/KG		
Sample II	)			
Tested	12/27/23			
Tester	MHUS			
Recorded				
INSTRUMENTS USED		INSTRUMENT		M103,T38
PREPARATION				*
Sample II	)			
Tested	12/27/23			
Tester	MHUS			
Recorded				
EXPIRATION DATE		MM/DD/YYYY		12/30/2025
		MM/DD/YYYY		
Sample II	)			
Tested	12/27/23			
Tester	MHUS			
Recorded				
GPL EDTA Disodium		3.72 GPL		3.72
	•	GPL		
Sample II	)			
Tested	12/27/23			
Tester	MHUS			
Recorded				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample II	)			
Tested	12/27/23			
Tester	MHUS			
Recorded				
recorded	12/2/120			

Page: 2

Date: 12/27/23 at 2:02 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Lot # 3122029 Made 12/27/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**