



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: HOU043F			
Aromatics by HPLC Quality Control Standard ASTM D6591			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 12/27/23			
Tester GWEB			
Recorded 12/27/23			
CALCULATED GRAMS CYCLOHEXANE/100ML	50.0000 GM/100ML	49.9900 - 50.0100	50.0002
Sample ID			
Tested 12/27/23			
Tester GWEB			
Recorded 12/27/23			
CALCULATED GRAMS O-XYLENE/100ML	15.0000 GM/100ML	14.9900 - 15.0100	14.9997
Sample ID			
Tested 12/27/23			
Tester GWEB			
Recorded 12/27/23			
CALCULATED GRAMS 1-METHYLNAPHTHALENE/100M	0.5000 GM/100ML	0.4800 - 0.5200	0.4998
Sample ID			
Tested 12/27/23			
Tester GWEB			
Recorded 12/27/23			
GM/100ML PHENANTHRENE (P1943)	0.3000 GRAMS	0.2600 - 0.3400	0.2997
Sample ID			
Tested 12/27/23			
Tester GWEB			
Recorded 12/27/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 12/27/23			
Tester GWEB			
Recorded 12/27/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M115,VF100-K9
Sample ID			
Tested 12/27/23			
Tester GWEB			
Recorded 12/27/23			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/24
Sample ID			
Tested 12/27/23			
Tester GWEB			
Recorded 12/27/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 12/27/23 at 5:08 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

1 YEAR
TIME

1 YEAR

Sample ID

Tested 12/27/23

Tester MMAT

Recorded 12/27/23

Lot # 3122711

Made 12/27/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager