

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

	NITROGEN	STANDARD 32.00% W/W	NITROGEN	
Test		Target/UOM	Range	Result
ACTUAL GRAMS WATER (GMS)		57.4732	57.4727 - 57.4737	57.4733
		GM		
Sample ID				
Tested	01/02/24			
Tester	MHAY			
Recorded	01/02/24			
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	01/02/24			
Tester	MHAY			
Recorded	01/02/24			
WEIGHT PERCENT NIT	ROGEN W/W	32.00	31.99 - 32.01	32.00
(CALCULATED)		% W/W		
Sample ID				
Tested	01/02/24			
Tester	MHAY			
Recorded	01/02/24			
ACTUAL GRAMS AMMO	NIUM	125.0000	124.9995 - 125.0005	125.0002
NITRATE (GMS)		GM		
Sample ID				
Tested	01/02/24			
Tester	MHAY			
Recorded	01/02/24			
ACTUAL GRAMS UREA		100.0000	99.9995 - 100.0005	99.9999
	(ullio)	GM		
Sample ID		Cim		
Tested	01/02/24			
Tester	MHAY			
Recorded	01/02/24			
EXPIRATION DATE	01/02/24	MM/DD/YY		01/30/25
		MM/DD/YYYY		01/30/23
O-male ID				
Sample ID	01/00/04			
Tested	01/02/24			
Tester	MHAY			
Recorded	01/02/24			
SHELF LIFE OF PRODU		SHELF LIFE		1 YEAR
. =		TIME		
Sample ID				
Tested	01/02/24			
Tester	MHAY			
Recorded	01/02/24			
Lot # 31229	913			
Made 01/02	2/24			



Page: 2 Date: 01/02/24 at 2:33 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

11hard SE

Quality Manager