



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: SPX475F | | | |
|--------------------------------------------------|-------------------------------------------|------------|--------------|
| NITROGEN STANDARD 1000 PPM AS N AQUEOUS SOLUTION | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON | YES | | YES |
| FILE | YES | | |
| Sample ID | | | |
| Tested 01/16/23 | | | |
| Tester MHUS | | | |
| Recorded 01/16/23 | | | |
| APPEARANCE AND COLOR | CLEAR WATER WHITE LIQUID APPEARANCE | | PASS |
| Sample ID | | | |
| Tested 01/16/23 | | | |
| Tester MHUS | | | |
| Recorded 01/16/23 | | | |
| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | | M108,VF1-Z16 |
| Sample ID | | | |
| Tested 01/16/23 | | | |
| Tester MHUS | | | |
| Recorded 01/16/23 | | | |
| EXPIRATION DATE | MM/DD/YY MM/DD/YYYY | | 03/18/23 |
| Sample ID | | | |
| Tested 01/16/23 | | | |
| Tester MHUS | | | |
| Recorded 01/16/23 | | | |
| PPM NITROGEN NO DECIAML PLACES | 1000 PPM | 985 - 1015 | 999 |
| Sample ID | | | |
| Tested 01/16/23 | | | |
| Tester MHUS | | | |
| Recorded 01/16/23 | | | |
| Lot # | 3011391 | | |
| Made | 01/16/23 | | |

Richard [Signature]

Quality Manager