

Page: 1

Date: 01/20/23 at 2:22 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	5851			
	METHYL	RED 0.1% W/V IN RGNT AL	COHOL	
Test		Target/UOM	Range	Result
PREPARED TO FORM	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	01/20/23			
Tester	ENOR			
Recorded	01/20/23			
W/V% METHYL RED		0.10	0.09 - 0.11	0.10
		W/V%		
Sample ID				
Tested	01/20/23			
Tester	ENOR			
Recorded	01/20/23			
APPEARANCE AND COLOR		CLEAR RED LIQUID		PASS
	-	APPEARANCE		
Sample ID		7.11.7.27.11.11.10.2		
Tested	01/20/23			
Tester	ENOR			
Recorded	01/20/23			
INSTRUMENTS USED		INSTRUMENT		M81,VF100-F1
PREPARATION	Domina	INOTITIONIE!		1001, 1110011
Sample ID				
Tested				
Tester	01/20/23 ENOR			
	01/20/23			
Recorded	01/20/23	MM/DD/YY		07/30/24
EXPIRATION DATE				07/30/24
		MM/DD/YYYY		
Sample ID				
Tested	01/20/23			
Tester	ENOR			
Recorded	01/20/23	40 MONTHS		40 140 17710
SHELF LIFE OF PROD	UCT	18 MONTHS TIME		18 MONTHS
Sample ID				
Tested	01/20/23			
Tester	ENOR			
Recorded	01/20/23			
<b>Lot #</b> 3010	321			
Made 01/2				



Page: 2

Date: 01/20/23 at 2:22 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**