Page: 1

Date: 01/17/23 at 2:17 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	CY395F			
Turbidity Blind QA Standard				
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	01/17/23			
Tester	MMAT			
Recorded	01/17/23			
TURBIDITY (NTU)		7.0	5.0 - 10.0	6.0
		NTU		
Sample ID				
•	01/17/23			
Tester	MMAT			
Recorded	01/17/23			
APPEARANCE AND COL		SLIGHTLY TURBID LIQUID		PASS
		APPEARANCE		
Sample ID				
•	01/17/23			
	MMAT			
	01/17/23			
INSTRUMENTS USED DU		INSTRUMENT		P49,VF1-Z2
PREPARATION				
Sample ID				
	01/17/23			
	MMAT			
	01/17/23			
EXPIRATION DATE	0.7.7720	MM/DD/YYYY		04/30/23
		MM/DD/YYYY		5 ,, 5 5, -5
Sample ID				
·	01/17/23			
	MMAT			
	01/17/23			
Lot # 301162				
Made 01/17/	' Z S			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I