Page: 1

Date: 01/20/23 at 9:21 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produ	uct: UCT126	5F		
		E STD FOR TSS 40 MG/L SOL	UTION	
Test		Target/UOM	Range	Result
PREPARED TO FOR	MULATION ON	YES		YES
FILE		YES		
Sample	ID			
Tested	01/20/23			
Tester	MMAT			
Recorde	ed 01/20/23			
ACTUAL PPM		40.00	39.50 - 40.50	39.98
		PPM		
Sample	ID			
Tested	01/20/23			
Tester	MMAT			
Recorde				
APPEARANCE AND COLOR		SLIGHTLY TURBID LIQUID		PASS
AIT EATIANCE AND	COLON	APPEARANCE		1 700
Sample	ID	ALL EXITATION		
Tested				
	01/20/23			
Tester	MMAT			
Recorde		INSTRUMENT		M21,VF4-N17
INSTRUMENTS USED DURING		INSTRUMENT		W21,VF4-N17
PREPARATION				
Sample				
Tested	01/20/23			
Tester	MMAT			
Recorde	ed 01/20/23			
EXPIRATION DATE		MM/DD/YY		07/30/23
		MM/DD/YYYY		
Sample	ID			
Tested	01/20/23			
Tester	MMAT			
Recorde	ed 01/20/23			
Lot # 301	L1937			
-0."	/20/23			
waue of	20/20			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I