



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: LY386F			
SULFUR STANDARD 10.0 PPM W/W IN ISOOCTANE			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 01/17/23			
Tester MMAT			
Recorded 01/17/23			
Calculated ppm Sulfur	10.00	9.80 - 10.20	10.00
	PPM		
Sample ID			
Tested 01/17/23			
Tester MMAT			
Recorded 01/17/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 01/17/23			
Tester MMAT			
Recorded 01/17/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P366,VF500-X34
Sample ID			
Tested 01/17/23			
Tester MMAT			
Recorded 01/17/23			
Lot #	3011773		
Made	01/17/23		

Quality Manager