



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: VUL029F			
	SODIUM HYDROXIDE 0.005 NORMAL		
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 01/19/23			
Tester JENC			
Recorded 01/19/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 01/19/23			
Tester JENC			
Recorded 01/19/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		01/19/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 01/19/23			
Tester JENC			
Recorded 01/19/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	23.9
SAMPLING DEGC	DEG C		
Sample ID			
Tested 01/19/23			
Tester JENC			
Recorded 01/19/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		01/19/23
	MM/DD/YYYY		
Sample ID			
Tested 01/19/23			
Tester JENC			
Recorded 01/19/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	23.9
DEG C	DEG C		
Sample ID			
Tested 01/19/23			
Tester JENC			
Recorded 01/19/23			
NORMALITY TO 4 DEC PLACES	0.0050	0.0047 - 0.0053	0.0049
	EQUIV/LITER		
Sample ID			
Tested 01/19/23			
Tester JENC			
Recorded 01/19/23			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 01/19/23			
Tester JENC			
Recorded 01/19/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Density	0.99847	0.99832 - 0.99861	0.99850
	GM/ML		
Sample ID			
Tested	01/19/23		
Tester	JENC		
Recorded	01/19/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		VF200-T7,VF4-I5,M55,BID0028
			0
Sample ID			
Tested	01/19/23		
Tester	JENC		
Recorded	01/19/23		
SHELF LIFE OF PRODUCT	6 MONTHS		6 MONTHS
	TIME		
Sample ID			
Tested	01/19/23		
Tester	JENC		
Recorded	01/19/23		
EXPIRATION DATE	MM/DD/YYYY		07/30/23
	MM/DD/YYYY		
Sample ID			
Tested	01/19/23		
Tester	JENC		
Recorded	01/19/23		
Lot #	3011847		
Made	01/19/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager