

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produ	uct: 6921			
-		TIVITY STANDARD 10.0 uMHO	S	
Test		Target/UOM	Range	Result
PREPARED TO FOR	RMULATION ON	YES	-	YES
FILE		YES		
Sample	ID			
Tested	01/04/24			
Tester	BPHI			
Record				
CONDUCTIVITY @ 25°C 1 DEC		10.0	9.9 - 10.1	10.0
PLACE		US/CM		
Sample				
Tested	01/04/24			
Tester	BPHI			
Record				
QCSP30:CONDUCTIVITY		YES		YES
STD&MEASUREME		QCSP30		
Sample				
Tested	01/04/24			
Tester	BPHI			
Record				
TRACEABLE TO N.I	.S.T. SRM 3191	YES		YES
		YES		
Sample				
Tested	01/04/24			
Tester	BPHI ad /04/04			
Record				DACC
APPEARANCE AND	COLOR	CLEAR WATER WHITE		PASS
		APPEARANCE		
Sample				
Tested	01/04/24			
Tester	BPHI ed 01/04/24			
Recorde		INSTRUMENT		M92,T234
PREPARATION	DURING	INSTROMENT		10152,1254
Sample	01/04/24			
Tested Tester	BPHI			
Record				
EXPIRATION DATE	ed 01/04/24	MM/DD/YY		04/30/24
		MM/DD/YYYY		01/00/21
Sample	חו			
Tested	01/04/24			
Tester	BPHI			
Record				
SHELF LIFE OF PR		SHELF LIFE		3 MONTHS
		TIME		
Sample	ID			
Tested	01/04/24			
Tester	BPHI			
Record	ed 01/04/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4010458
Made	01/04/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard St

Quality Manager