

Made 01/27/24

Page: 1

Date: 01/27/24 at 7:05 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

		- Odstomer i		
Packaged Product:				
r	CONDUC'	TIVITY STANDARD 5.0 uMHOS		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	01/27/24			
Tester	BPHI			
Recorded	01/27/24			
CONDUCTIVITY		5.0	4.8 - 5.2	5.0
		UMHOS		
Sample ID				
Tested	01/27/24			
Tester	BPHI			
Recorded	01/27/24			
QCSP30:CONDUCTIVIT	Υ	YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
Tested	01/27/24			
Tester	BPHI			
Recorded	01/27/24			
TRACEABLE TO N.I.S.T	. SRM 3191	YES		YES
		YES		
Sample ID				
Tested	01/27/24			
Tester	BPHI			
Recorded	01/27/24			
APPEARANCE AND CO	LOR	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	01/27/24			
Tester	BPHI			
Recorded	01/27/24			
INSTRUMENTS USED D		INSTRUMENT		M92,T234
PREPARATION	. •			
Sample ID				
Tested	01/27/24			
Tester	BPHI			
Recorded	01/27/24			
EXPIRATION DATE	v	MM/DD/YY		04/30/24
		MM/DD/YYYY		
Sample ID				
Tested	01/27/24			
Tester	BPHI			
Recorded	01/27/24			
Lot # 40126				
LOT # 40120				



Page: 2

Date: 01/27/24 at 7:05 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager