

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produc	spx303	ਸ ਸ		
Fackageu Flouut		NITRATE 0.02 MOLAR AQUEC		
	CUPRIC			
Test		Target/UOM	Range	Result
PREPARED TO FORM	IULATION ON	YES		YES
FILE		YES		
Sample I	D			
Tested	01/23/24			
Tester	MHAY			
Recorded				
APPEARANCE AND COLOR		CLEAR BLUE LIQUID		PASS
		APPEARANCE		
Sample I	D			
Tested	01/23/24			
Tester	MHAY			
Recorded	01/23/24			
INSTRUMENTS USED	DURING	INSTRUMENT		M108,GC
PREPARATION				
Sample I	D			
Tested	01/23/24			
Tester	MHAY			
Recorded	01/23/24			
EXPIRATION DATE		MM/DD/YY		01/30/25
		MM/DD/YYYY		
Sample I	D			
Tested	01/23/24			
Tester	MHAY			
Recorded	01/23/24			
MOLARITY CAL AS CU(NO2)3 X		0.020 MOLARITY		0.020
3H20		MOLARITY (M/L)		
Sample I	D			
Tested	01/23/24			
Tester	MHAY			
Recorded				
Lot # 4012	2297			
	23/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA