02/01/24

Made

Page: 1

Date: 02/01/24 at 5:33 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	1792			
-	CHLORINE	STANDARD 100 PPM AS	CL2 CERTIFIED SOLUTION	
Test		Target/UOM	Range	Result
SHELF LIFE OF PRODUCT		2 MONTHS		2 MONTHS
		TIME		
Sample ID				
Tested (02/01/24			
Tester S	SHAR			
Recorded (02/01/24			
Actual Density		0.99951	0.99884 - 1.00018	0.99917
		GM/ML		
Sample ID				
	02/01/24			
	SHAR			
Recorded (02/01/24			
INSTRUMENTS USED DU	RING	INSTRUMENT		M112,VF12-4,M55,BID079,PID
PREPARATION				P168
Sample ID				
-	02/01/24			
	SHAR			
	02/01/24			
EXPIRATION DATE		MM/DD/YY		04/05/24
		MM/DD/YYYY		
Sample ID				
	02/01/24			
	SHAR			
	02/01/24			
PPM CHLORIDE		100.00	98.00 - 102.00	99.95
		PPM		
Sample ID				
	02/01/24			
	SHAR			
	02/01/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland St