Page: 1

Date: 01/30/24 at 5:59 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

<b>Package</b>	d Product:	0370			
		AMMONIA	STOCK STANDARD 500 pp	m AS NH3	
Test			Target/UOM	Range	Result
PREPAREI	TO FORMUL	ATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	01/30/24			
	Tester	AROS			
	Recorded	01/30/24			
CERTIFIED	PPM NH3		500.0	490.0 - 510.0	500.8
			PPM		
	Sample ID				
		01/30/24			
		AROS			
	Recorded	01/30/24			
APPEARANCE AND COLOR		OR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	01/30/24			
	Tester	AROS			
	Recorded	01/30/24			
INSTRUMENTS USED DURING		IRING	INSTRUMENT		M115,GC,VF4-N31
PREPARAT	TION				
	Sample ID				
	•	01/30/24			
		AROS			
	Recorded	01/30/24			
EXPIRATION DATE			MM/DD/YYYY		05/30/2024
			MM/DD/YYYY		
	Sample ID				
	•	01/30/24			
		AROS			
		01/30/24			
	t# 401300	) 8			
Mad	ue 01/30/	<u>_</u> ¬			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Parland St