

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

	0 - 0 4			1
Packaged Product:	9594			
	TOC STANDARD	10,000 PPM (KHP)		
Test	Ta	rget/UOM	Range	Result
PREPARED TO FORMUL		YES		YES
FILE		YES		
Sample ID				
Tested	01/24/24			
Tester	DTRA			
Recorded	01/24/24			
APPEARANCE AND COL	OR CL	EAR LIQUID		PASS
	AP	PEARANCE		
Sample ID				
Tested	01/24/24			
Tester	DTRA			
Recorded	01/24/24			
INSTRUMENTS USED DU	JRING IN	STRUMENT	Ν	//81,VF2-A6
PREPARATION				
Sample ID				
Tested	01/24/24			
Tester	DTRA			
Recorded	01/24/24			
EXPIRATION DATE	Ν	MM/DD/YY		04/30/24
	M	M/DD/YYYY		
Sample ID				
Tested	01/24/24			
Tester	DTRA			
Recorded	01/24/24			
Lot # 401221	133			
Made 01/24/	/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stol

Quality Manager