

Page: 1

Date: 01/30/24 at 5:59 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: 0372			
		STANDARD 100 PPM AS NI		
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID	)			
Tested	01/30/24			
Tester	AROS			
Recorded	01/30/24			
CERTIFIED PPM NH3		100.0	96.0 - 104.0	100.8
		PPM		
Sample ID	)			
Tested	01/30/24			
Tester	AROS			
Recorded	01/30/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID	)			
Tested	01/30/24			
Tester	AROS			
Recorded				
INSTRUMENTS USED		INSTRUMENT		M115,GC,VF6-W5
PREPARATION				, ,
Sample ID	)			
Tested	01/30/24			
Tester	AROS			
Recorded	01/30/24			
EXPIRATION DATE	01/00/21	MM/DD/YYYY		05/30/2024
EXI II CATION DATE		MM/DD/YYYY		00/03/2021
Sample ID	1	, <i>DD</i> , 1.11		
Tested	01/30/24			
Tester	01/30/24 AROS			
Recorded				
SHELF LIFE OF PROD		3 MONTHS		3 MONTHS
SHELF LIFE OF FROD	,001	TIME		3 WONTIS
Sample ID	)			
Tested	01/30/24			
Tester	AROS			
Recorded				
<b>Lot #</b> 4013				
<b>Made</b> 01/3				
illudo , c	•			



Page: 2

Date: 01/30/24 at 5:59 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**