

Page: 1

Date: 01/22/24 at 4:43 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: CG034F			
	CHLORID	E STANDARD 1.0 MG/KG IN	1 P-XYLENE	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	01/22/24			
Tester	MHAY			
Recorded	01/22/24			
PPM CHLORIDE		1.00	0.90 - 1.10	1.00
		PPM		
Sample ID				
Tested	01/22/24			
Tester	MHAY			
Recorded	01/22/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	01/22/24			
Tester	MHAY			
Recorded	01/22/24			
INSTRUMENTS USED I		INSTRUMENT		P89,VF1-Z13
PREPARATION				
Sample ID				
Tested	01/22/24			
Tester	MHAY			
Recorded	01/22/24			
EXPIRATION DATE	01/22/24	MM/DD/YY		07/30/24
EXPINATION DATE		MM/DD/YYYY		01750/24
Sample ID		IVIIVII DDI I I I I		
Tested	01/22/24			
Tested	01/22/24 MHAY			
Recorded	01/22/24			
SHELF LIFE OF PRODI		SHELF LIFE		6 MONTHS
SHELF LIFE OF PRODU	UC I	TIME		O INIOINI D
Sample ID				
Tested	01/22/24			
Tester	MHAY			
Recorded	01/22/24			
Lot # 4012296				
Made 01/22/24				



Page: 2

Date: 01/22/24 at 4:43 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**