Page: 1

Date: 01/10/24 at 11:11 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

	Color S	olor Solution for EZ1012/EZ2500 Total Cyanide Analyzer		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES	-	YES
FILE		YES		
Sample ID				
Tested	01/10/24			
Tester	BPHI			
Recorded	01/10/24			
PH OF PRODUCT/SOLUTION		5.200	5.160 - 5.240	5.173
		PH		
Sample ID				
Tested	01/10/24			
Tester	BPHI			
Recorded	01/10/24			
APPEARANCE AND COLOR		APPEARANCE		PASS
Sample ID				
Tested	01/10/24			
Tester	BPHI			
Recorded	01/10/24			
INSTRUMENTS USED I	DURING	INSTRUMENT		M103,EF1-PLD
PREPARATION				
Sample ID				
Tested	01/10/24			
Tester	BPHI			
Recorded	01/10/24			
EXPIRATION DATE		MM/DD/YY		03/10/24
		MM/DD/YYYY		
Sample ID				
Tested	01/10/24			
Tester	BPHI			
Recorded	01/10/24			
Lot # 4010	542			
Made 01/1				
wade 01/1	0 / 4 7			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hichard Stol