

Page: 1

Date: 01/11/24 at 5:48 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product				
	BORIC A	ACID 1.7% W/V pH ADJUSTE	D TO 4.6	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	01/11/24			
Tester	BPHI			
Recorded	01/11/24			
PH OF PRODUCT/SOLUTION		4.600	4.550 - 4.650	4.599
		PH		
Sample ID				
Tested	01/11/24			
Tester	BPHI			
Recorded	01/11/24			
CALCULATED W/V% B		1.70	1.69 - 1.71	1.70
		W/V%		
Sample ID				
Tested	01/11/24			
Tester	BPHI			
Recorded	01/11/24			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
A I LAIGHTOL AND GO	LOR	APPEARANCE		17.66
Sample ID		ALLENVINOE		
Tested	01/11/24			
Tester	BPHI			
Recorded	01/11/24			
		INSTRUMENT		M103,VF12-1,M106
INSTRUMENTS USED DURING		INSTRUMENT		W1103, V F 12-1, W1100
PREPARATION				
Sample ID				
Tested	01/11/24			
Tester	BPHI			
Recorded	01/11/24	MAUDDOOGG		04/00/0005
EXPIRATION DATE		MM/DD/YYYY MM/DD/YYYY		01/30/2025
Sample ID				
Sample ID Tested	04/44/04			
Tested	01/11/24 BPHI			
Recorded	01/11/24			
Lot # 4010956				
Made 01/11/24				



Page: 2

Date: 01/11/24 at 5:48 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager