

Page: 1

Date: 01/16/24 at 3:05 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Package	d Product	LY346F			
		SODIUM	STANDARD 400 PPM IN 2%	HCL	
Test			Target/UOM	Range	Result
	TO FORMU	ILATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	01/16/24			
	Tester	MHAY			
	Recorded	01/16/24			
PPM NA			400.00	399.00 - 401.00	399.93
			PPM		
	Sample ID				
	Tested	01/16/24			
	Tester	MHAY			
	Recorded	01/16/24			
APPEARANCE AND COLOR			CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	01/16/24			
	Tester	MHAY			
	Recorded	01/16/24			
INSTRUME	NTS USED I		INSTRUMENT		M108,GC,VF1-Z25
PREPARAT					
	Sample ID				
	Tested	01/16/24			
	Tester	MHAY			
	Recorded	01/16/24			
SHFLF LIF	E OF PRODU		6 MONTHS		6 MONTHS
011221 211	_ 0 0.		TIME		c me.tine
	Sample ID		· <u> </u>		
	Tested	01/16/24			
	Tester	MHAY			
	Recorded	01/16/24			
EXPIRATION		51/10/LT	MM/DD/YYYY		07/30/2024
EXI IIIAIIC	N DAIL		MM/DD/YYYY		07/00/2021
	Sample ID				
	Tested	01/16/24			
	Tester	MHAY			
	Recorded	01/16/24			
Lo	t# 40116				
Made 01/16/24					
ivia	uc -/ - \				



Page: 2

Date: 01/16/24 at 3:05 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager