

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	t: 265-01F			
		NITRATE 0.0282 NORMAL	ANTI-FREEZE	
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES	~	YES
FILE		YES		
Sample ID)			
Tested	02/01/24			
Tester	JENC			
Recorded	02/01/24			
NORMALITY TO 4 DEC PLACES		0.0282	0.0280 - 0.0284	0.0281
		EQUIV/LITER		
Sample ID)			
Tested	02/01/24			
Tester	JENC			
Recorded	02/01/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID)			
Tested	02/01/24			
Tester	JENC			
Recorded	02/01/24			
INSTRUMENTS USED DURING		INSTRUMENT		M103,VF6-W9,GC,T-220
PREPARATION				
Sample ID)			
Tested	02/01/24			
Tester	JENC			
Recorded	02/01/24			
EXPIRATION DATE		MM/DD/YYYY		02/28/2026
		MM/DD/YYYY		
Sample ID)			
Tested	02/01/24			
Tester	JENC			
Recorded				
Lot # 4012	639			
Made 02/0	1/24			
	_,			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA