

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	FOX047F			
		n Permanganate 0.005	Normal NIST Traceable	
Test		Target/UOM	Range	Result
PREPARED TO FORM	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	01/23/24			
Tester	MHUS			
Recorded	01/23/24			
CALCULATED NORMA		0.0050	0.0049 - 0.0051	0.0050
		EQUIV/LITER		
Sample ID				
Tested	01/23/24			
Tester	MHUS			
Recorded	01/23/24			
APPEARANCE AND COLOR		PURPLE LIQID		PASS
		APPEARANCE		
Sample ID				
Tested	01/23/24			
Tester	MHUS			
Recorded	01/23/24			
		INSTRUMENT		VF1-Z32,T223
PREPARATION				,
Sample ID				
Tested	01/23/24			
Tester	MHUS			
Recorded	01/23/24			
EXPIRATION DATE		MM/DD/YYYY		07/30/2024
· · · · · · · · · · · · · · · · · · ·		MM/DD/YYYY		
Sample ID				
Tested	01/23/24			
Tester	MHUS			
Recorded	01/23/24			
Lot # 4012	261			
Made 01/2				
inddo				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA