

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	• DC896F			
	•	ard 20.0 ppm Certifi	ed Solution	
Test		Target/UOM	Range	Result
PREPARED TO FORMU		YES	Kange	YES
FILE		YES		120
Sample ID		120		
Tested	01/30/24			
Tester	DTRA			
Recorded	01/30/24			
Certified ppm TOC	01/30/24	20.00	19.60 - 20.40	19.99
		PPM (MG/L)	19.00 - 20.40	19.99
Complet D				
Sample ID	04/00/04			
Tested	01/30/24			
Tester	DTRA			
	01/30/24			PASS
APPEARANCE AND CO	LOR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	01/30/24			
Tester	DTRA			
Recorded	01/30/24			
INSTRUMENTS USED DURING		INSTRUMENT		P175,VF5-L14
PREPARATION				
Sample ID				
Tested	01/30/24			
Tester	DTRA			
Recorded	01/30/24			
EXPIRATION DATE		MM/DD/YYYY		05/30/2024
		MM/DD/YYYY		
Sample ID				
Tested	01/30/24			
Tester	DTRA			
Recorded	01/30/24			
Lot # 40126	502			
Made 01/30				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA