

# Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

#### Customer PO # :

Packaged Product: 8430			
-	ITRATE 0.1 NORMAL NIST TRACEABLE SOLUTION		
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 01/18/24			
Tester DTRA			
Recorded 01/18/24			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 01/18/24			
Tester DTRA			
Recorded 01/18/24			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		01/17/2024
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 01/18/24			
Tester DTRA			
Recorded 01/18/24			
	25.0	10.0 - 35.0	20.0
SAMPLING DEGC	DEG C		20.0
Sample ID			
Tested 01/18/24			
Tester DTRA			
Recorded 01/18/24	MM/DD/YYYY		01/17/2024
DATE OF TESTING OF SAMPLE			01/17/2024
	MM/DD/YYYY		
Sample ID			
Tested 01/18/24			
Tester DTRA			
Recorded 01/18/24			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0
DEG C	DEG C		
Sample ID			
Tested 01/18/24			
Tester DTRA			
Recorded 01/18/24			
Actual Normality	0.1000	0.0995 - 0.1005	0.1001
	EQ/L		
Sample ID			
Tested 01/18/24			
Tester DTRA			
Recorded 01/18/24			
QCSP53: SILVER NITRATE STD ON	YES		YES
AUTO-TITRA	METHOD		
Sample ID			
Tested 01/18/24			
Tester DTRA			
Recorded 01/18/24			



## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

#### Customer PO # :

	Cus		
Actual Density	1.01229 GM/ML	1.00868 - 1.01590	1.01249
Sample ID			
Tested 01/18/2	4		
Tester DTRA	•		
Recorded 01/18/2	4		
VERIFICATION BY QCSP78	PASS		PASS
	PASS		1760
Sample ID			
Tested 01/18/2	1		
Tester DTRA	-		
Recorded 01/18/2	4		
N.I.S.T. Traceable to SRM 919 N			YES
N.I.S.T. HACEADIE LO SKW 919 N	YES		TES
Sample ID	TEO		
Sample ID Tested 01/18/2	<i>A</i>		
	4		
Tester DTRA Recorded 01/18/2	<i>A</i>		
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
AFFEARANCE AND COLOR	LIQUID		1 400
	APPEARANCE		
Commis ID	AFFEARANCE		
Sample ID			
Tested 01/18/2	4		
Tester DTRA			
Recorded 01/18/2			
INSTRUMENTS USED DURING	INSTRUMENT		M114,T211,M55,TID53294
PREPARATION			
Sample ID			
Tested 01/18/2	4		
Tester DTRA			
Recorded 01/18/2			04/00/00
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		01/30/26
Sample ID			
Tested 01/18/2	4		
Tester DTRA			
Recorded 01/18/2	4		
NORMALITY USING	0.1000	0.0985 - 0.1015	0.1001
AUTO-TITRATOR	EQUIV/LITER		
Sample ID			
Tested 01/18/2	4		
Tester DTRA			
Recorded 01/18/2	4		
SHELF LIFE OF PRODUCT	SHELF LIFE		2 YEARS
	TIME		
Sample ID			
Tested 01/18/2	4		
Tester DTRA			
Recorded 01/18/2	4		
Lot # 4010960			
Made 01/18/24			



Page: 3 Date: 01/18/24 at 11:07 AM

### Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1 chard SE

**Quality Manager**