

Page: 1

Date: 01/18/24 at 11:17 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:				
_	TC/TN MI	EXED STANDARD (200 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	01/18/24			
Tester	GWEB			
Recorded	01/18/24			
CERTIFIED TOTAL ORGANIC		200.0	195.0 - 205.0	199.5
CARBON (PPM)		PPM		
Sample ID				
Tested	01/18/24			
Tester	GWEB			
Recorded	01/18/24			
CALCULATED PPM NIT	ROGEN	1000	995 - 1005	1000
		PPM		
Sample ID				
Tested	01/18/24			
Tester	GWEB			
Recorded	01/18/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
,		APPEARANCE		
Sample ID		711 1 2711 0 11 0 2		
Tested	01/18/24			
Tester	GWEB			
Recorded	01/18/24			
		INSTRUMENT		M108,VF5-E5
INSTRUMENTS USED DURING		INSTRUMENT		W100, V1 3-L3
PREPARATION				
Sample ID	0.1.1.0.10.1			
Tested	01/18/24			
Tester	GWEB			
Recorded	01/18/24	MANDENAGG		00/00/0004
EXPIRATION DATE		MM/DD/YYYY MM/DD/YYYY		03/23/2024
Sample ID				
Tested	01/18/24			
Tester	GWEB			
Recorded	01/18/24			
Lot # 40117				
Made 01/18				
111440				



Page: 2

Date: 01/18/24 at 11:17 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager