Page: 1

Date: 01/03/24 at 3:18 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	ct: UCT12	6F		
		E STD FOR TSS 40 MG/L SOLU	JTION	
Test		Target/UOM	Range	Result
PREPARED TO FORM	MULATION ON	YES		YES
FILE		YES		
Sample I	D			
Tested	01/03/24			
Tester	AROS			
Recorded	d 01/03/24			
ACTUAL PPM		40.00	39.50 - 40.50	40.00
		PPM		
Sample I	D			
Tested	01/03/24			
Tester	AROS			
Recorded	d 01/03/24			
APPEARANCE AND COLOR		SLIGHTLY TURBID LIQUID		PASS
		APPEARANCE		
Sample I	D			
Tested	01/03/24			
Tester	AROS			
Recorded				
INSTRUMENTS USED DURING		INSTRUMENT		M108,VF4-N26
PREPARATION				
Sample I	D			
Tested	01/03/24			
Tester	AROS			
Recorded				
EXPIRATION DATE		MM/DD/YY		07/30/24
		MM/DD/YYYY		
Sample I	D			
Tested	01/03/24			
Tester	AROS			
Recorded				
Lot # 401	0305			
	03/24			
Widde •=	•			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I