Page: 1

Date: 01/17/24 at 9:23 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: DC893F			
		UM IODATE 0.0025 NORMA	L	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES	<u> </u>	YES
FILE		YES		
Sample II)			
Tested	01/17/24			
Tester	JENC			
Recorded	01/17/24			
NORMALITY TO 4 DEC PLACES		0.0025	0.0023 - 0.0027	0.0025
		EQUIV/LITER		
Sample II				
Tested	01/17/24			
Tester	JENC			
Recorded				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample II)	74 7 274 6 44 62		
Tested	01/17/24			
Tester	JENC			
Recorded				
INSTRUMENTS USED		INSTRUMENT		M91,M113,M112,T9
PREPARATION		INCTROMENT		1010 1,1011 10,1011 12,10
Sample II Tested				
	01/17/24			
Tester	JENC			
Recorded EXPIRATION DATE	01/17/24	MM/DD/YY		01/30/25
EXPIRATION DATE				01/30/25
		MM/DD/YYYY		
Sample II				
Tested	01/17/24			
Tester	JENC			
Recorded				
Lot # 4011005				
Made 01/1	7/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Parland St