01/11/24

Made

Page: 1

Date: 01/11/24 at 4:12 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	ct: 6492			
		GEN STANDARD 500 PPM AS N		
Test		Target/UOM	Range	Result
PREPARED TO FORM	MULATION	YES		YES
		YES		
Sample I	ID			
Tested	01/11/24			
Tester	AROS			
Recorde				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample I	ID			
Tested	01/11/24			
Tester	AROS			
Recorde	d 01/11/24			
INSTRUMENTS USED DURING		LAB EQUIPMENT USED		M108,GC,VF1-Z39
PREPARATION		INSTRUMENT		
Sample I	ID			
Tested	01/11/24			
Tester	AROS			
Recorde	d 01/11/24			
CERTIFIED PPM NITROGEN:		500.0	485.0 - 515.0	502.8
		PPM		
Sample I	ID			
Tested	01/11/24			
Tester	AROS			
Recorde	d 01/11/24			
EXPIRATION DATE		MM/DD/YYYY		04/30/2024
		MM/DD/YYYY		
Sample I	ID			
Tested	01/11/24			
Tester	AROS			
Recorde	d 01/11/24			
Lot # 401	1142			
	11/04			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Pareland St