

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Package	d Product:	OXY517F			
. uonugo			Copper and Chromium in	0.15% HNO3 solution	
Test			Target/UOM	Range	Result
	TO FORMU	ATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	01/19/24			
	Tester	AROS			
	Recorded	01/19/24			
PPM CU			5.00	4.95 - 5.05	5.00
			PPM		
	Sample ID				
	Tested	01/19/24			
	Tester	AROS			
	Recorded	01/19/24			
PPM CHRO	MIUM		5.00	4.95 - 5.05	5.00
			PPM		
	Sample ID				
	Tested	01/19/24			
	Tester	AROS			
	Recorded	01/19/24			5400
APPEARAN	ICE AND COL	OR	CLEAR		PASS
			APPEARANCE		
	Sample ID	01/10/04			
	Tested	01/19/24			
	Tester Recorded	AROS 01/19/24			
			INSTRUMENT		P41,P216,GC,VF1-Z35
INSTRUMENTS USED DURING PREPARATION		UNING	INSTRUMENT		141,1210,00,011-200
	Sample ID				
	Tested	01/19/24			
	Tester	AROS			
	Recorded	01/19/24			
Actual Den			0.99925	0.99796 - 1.00053	0.99888
	,		GM/ML		
	Sample ID				
	Tested	01/19/24			
	Tester	AROS			
	Recorded	01/19/24			
EXPIRATIO	N DATE		MM/DD/YY		01/30/25
			MM/DD/YYYY		
	Sample ID				
	Tested	01/19/24			
	Tester	AROS			
	Recorded	01/19/24			
SHELF LIFE OF PRODU		СТ	1 YEAR TIME		1 YEAR
	Sample ID				
	Tested	01/19/24			
	Tester	AROS			
	Recorded	01/19/24			



## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4011944
Made	01/19/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard St

**Quality Manager**