

Page: 1

Date: 01/27/24 at 6:59 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc		/ Managara 10 and	San DEC March	
T4	Arsenic	/ Manganese 10 ppm ea		Dogulf
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID)			
Tested	01/26/24			
Tester	MMAT			
Recorded	01/26/24			
CALCULATED PPM ARSENIC		10.00	9.95 - 10.05	10.00
		PPM / MG/L/L		
Sample ID)			
Tested	01/26/24			
Tester	MMAT			
Recorded	01/26/24			
CALCULATED PPM MA	ANGANESE	10.00	9.95 - 10.05	10.00
		PPM (UG/ML)		
Sample ID)	,		
Tested	01/26/24			
Tester	MMAT			
Recorded	01/26/24			
APPEARANCE AND CO		CLEAR LIQUID		PASS
AI I LAINANGE AND O	o Lor	APPEARANCE		17.66
Sample ID		ALLEAVANOL		
Tested				
	01/26/24			
Tester	MMAT			
Recorded	01/26/24	INICTOLIMENT		D240 D04 VE4 722
INSTRUMENTS USED DURING		INSTRUMENT		P340,P94,VF1-Z33
PREPARATION				
Sample ID				
Tested	01/26/24			
Tester	MMAT			
Recorded	01/26/24			
EXPIRATION DATE		MM/DD/YYYY		07/30/2024
		MM/DD/YYYY		
Sample ID)			
Tested	01/26/24			
Tester	MMAT			
Recorded	01/26/24			
Lot # 4012	521			
Made 01/2				
iliado/ =				



Page: 2

Date: 01/27/24 at 6:59 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager