02/01/24

Made

Page: 1

Date: 02/01/24 at 1:07 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

		Fluoride Standard 0.25 mg/L for EZ3000 Fluoride Analyzer				
Test			Target/UOM	Range	Result	
PREPARE	D TO FORMUL	ATION ON	YES		YES	
FILE			YES			
	Sample ID					
	Tested	02/01/24				
	Tester	MHAY				
	Recorded	02/01/24				
APPEARANCE AND COLOR		.OR	CLEAR WATER WHITE		PASS	
			APPEARANCE			
	Sample ID					
	Tested	02/01/24				
	Tester	MHAY				
	Recorded	02/01/24				
CALCULA [*]	TED PPM (MG/	/ L)	0.25	0.24 - 0.26	0.25	
FLUORIDE			MG/L			
	Sample ID					
	Tested	02/01/24				
	Tester	MHAY				
	Recorded	02/01/24				
INSTRUMENTS USED DURING		URING	INSTRUMENT		P89,VF4-N10	
PREPARA	TION					
	Sample ID					
	Tested	02/01/24				
	Tester	MHAY				
	Recorded	02/01/24				
EXPIRATION DATE			MM/DD/YY		08/30/24	
			MM/DD/YYYY			
	Sample ID					
		02/01/24				
	Tester	MHAY				
	Recorded	02/01/24				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

lauland I