

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Deelegrad Dre	duct: 9576			
Packaged Pro	4400			
	TOC STANDARD 3000 PPM			
Test		Target/UOM	Range	Result
PREPARED TO F	ORMULATION ON	YES		YES
FILE		YES		
Sam	ole ID			
Teste	ed 01/19/24			
Teste	er MHAY			
Reco	rded 01/19/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sam	ble ID			
Teste	ed 01/19/24			
Teste	er MHAY			
Reco	rded 01/19/24			
INSTRUMENTS USED DURING		LAB EQUIPMENT		M115,GC,VF2-A6
PREPARATION		INSTRUMENT		
Sam	ole ID			
Teste	ed 01/19/24			
Teste	er MHAY			
Reco				
ACTUAL PPM CERTIFIED TOC		3000	2800 - 3200	3046
VALUE		PPM		
Sam	ble ID			
Teste	ed 01/19/24			
Teste				
Reco				
EXPIRATION DATE		MM/DD/YYYY		04/30/2024
		MM/DD/YYYY		
Sam				
Teste				
Teste				
Reco	rded 01/19/24			
Lot # 4	011807			
	1/19/24			
maac	· ·			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA