Page: 1

Date: 01/30/24 at 4:15 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Prod	luct: LY161F			
_		SIUM STANDARD 0.5 PPM IN 3	METHANOL	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION		YES		YES
		YES		
Sampl	le ID			
Tested	01/30/24			
Tester	AROS			
Recor	ded 01/30/24			
CALCULATED PPM POTASSIUM		0.50	0.40 - 0.60	0.50
		MG/L		
Sampl	le ID			
Tested				
Tester				
Recor				
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		APPEARANCE		
Sampl	le ID			
Tested				
Tester				
Recor				
INSTRUMENTS USED DURING		LAB EQUIPMENT		P396,VF500-X42
PREPARATION		INSTRUMENT		,
Sampl	le ID			
Tested				
Tester	0 1/00/2 1			
Recor				
EXPIRATION DATE		MM/DD/YYYY		03/05/2024
	-	MM/DD/YYYY		
Sampl	le ID	,22,		
Tested				
Tester	0 1/00/2 1			
Recor				
Lot # 40				
Made 01	1/30/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I