

Page: 1

Date: 02/06/24 at 8:07 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: T3501F			
_		UTYLAMMONIUM HYDROXIDE 1	1.0 MOLAR/MeOH	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	02/06/24			
Tester	MHUS			
Recorded	02/06/24			
ACTUAL MOLARITY		1.0000	0.9900 - 1.0100	0.9980
		MOLARITY (M/L)		
Sample ID		,		
Tested	02/06/24			
Tester	MHUS			
Recorded	02/06/24			
Actual Density	0L/00/L-	0.85266	0.85112 - 0.85420	0.85178
Actual Denoity		GM/ML	0.00112 0.00420	0.00170
Sample ID		GIVI/IVIL		
Tested	00/00/04			
	02/06/24			
Tester	MHUS			
Recorded	02/06/24	DALE VELLOW LIQUID		PASS
APPEARANCE AND CO	LOR	PALE YELLOW LIQUID		PA55
		APPEARANCE		
Sample ID				
Tested	02/06/24			
Tester	MHUS			
Recorded	02/06/24			
INSTRUMENTS USED [DURING	INSTRUMENT		M87,VF5-E5
PREPARATION				
Sample ID				
Tested	02/06/24			
Tester	MHUS			
Recorded	02/06/24			
EXPIRATION DATE		MM/DD/YYYY		08/30/2025
		MM/DD/YYYY		
Sample ID				
Tested	02/06/24			
Tester	MHUS			
Recorded	02/06/24			
Lot # 4020139				
Made 02/06	0/24			



Page: 2

Date: 02/06/24 at 8:07 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager