

Page: 1

Date: 02/06/24 at 1:12 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

		Customer PO # :	
Packaged Product:	6919		
	CONDUCTIVITY STANDARD 10	00 uMHOS	
Test	Target/UOM	Range	Result
PREPARED TO FORMULAT	ION ON YES	Hange	YES
FILE	YES		
Sample ID			
	06/24		
Tester JR			
	06/24		
CONDUCTIVITY @25 C	100	99 - 101	100
	US/CM		
Sample ID			
·	06/24		
Tester JR			
	06/24		
QCSP30:CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
· ·	06/24		
Tester JR	10		
Recorded 02/	06/24		
TRACEABLE TO N.I.S.T. SR	M 3191 YES		YES
	YES		
Sample ID			
	06/24		
Tester JR	0		
Recorded 02/	06/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 02/	06/24		
Tester JR	10		
Recorded 02/	06/24		
INSTRUMENTS USED DURI	NG INSTRUMENT		M92,T-229
PREPARATION			
Sample ID			
Tested 02/	06/24		
Tester JR	10		
	06/24		
EXPIRATION DATE	MM/DD/YY		02/28/25
	MM/DD/YYYY		
Sample ID			
	06/24		
Tester JR			
	06/24		
SHELF LIFE OF PRODUCT	1 YEAR		1 YEAR
	TIME		
Sample ID			
	06/24		
Tester JR			
Recorded 02/	06/24		

Page: 2

Date: 02/06/24 at 1:12 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Lot # 4020620 Made 02/06/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager