02/05/24

Made

Page: 1

Date: 02/05/24 at 8:37 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: $^{ m LY44}$	13F		
	A 5.0 ppm w/v Standard		
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION O			YES
TLE	YES		
Sample ID			
Tested 02/05/24			
Tester MMAT			
Recorded 02/05/24			
CALCULATED PPM MDEA	5.00	4.95 - 5.05	5.00
	PPM		
Sample ID			
Tested 02/05/24			
Tester MMAT			
Recorded 02/05/24			
APPEARANCE AND COLOR	CLEAR LIQUID/ODORLESS		PASS
	APPEARANCE		
Sample ID			
Tested 02/05/24			
Tester MMAT			
Recorded 02/05/24			
NSTRUMENTS USED DURING	INSTRUMENT		P399,VF500-X44
PREPARATION			
Sample ID			
Tested 02/05/24			
Tester MMAT			
Recorded 02/05/24			
EXPIRATION DATE	MM/DD/YY		04/10/24
	MM/DD/YYYY		
Sample ID			
Tested 02/05/24			
Tester MMAT			
Recorded 02/05/24			
Lot # 4020529			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Pareland St