

**Made** 02/09/24

Page: 1

Date: 02/09/24 at 2:57 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results
Customer PO #:

Package	ed Product	: 0312			
		ALKALIN	ITY STANDARD 1000 PPM		
Test			Target/UOM	Range	Result
	ED TO FORMU	JLATION ON	YES	<u> </u>	YES
FILE			YES		
	Sample ID				
	Tested	02/09/24			
	Tester	MHAY			
	Recorded	02/09/24			
PPM ALKALINITY AS CACO3		ACO3	1000.0	995.0 - 1005.0	999.3
			PPM		
	Sample ID				
	Tested	02/09/24			
	Tester	MHAY			
	Recorded	02/09/24			
APPEARANCE AND COLOR		LOR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	02/09/24			
	Tester	MHAY			
	Recorded	02/09/24			
INSTRUM	IENTS USED D	DURING	INSTRUMENT		M115,VF500-X49
PREPAR/	ATION				
	Sample ID				
	Tested	02/09/24			
	Tester	MHAY			
	Recorded	02/09/24			
SHELF LIFE OF PRODUCT		JCT	3 MONTHS		3 MONTHS
			TIME		
	Sample ID				
	Tested	02/09/24			
	Tester	MHAY			
	Recorded	02/09/24			
EXPIRATION DATE			MM/DD/YYYY		05/30/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	02/09/24			
	Tester	MHAY			
	Recorded	02/09/24			
1	ot # 40209	933			
	•• II				



Page: 2

Date: 02/09/24 at 2:57 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**