



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: 9766-10LHP | | | |
|--------------------------------------|----------------------|-------|--------------|
| WATER DISTILLED ASTM TYPE II (10LHP) | | | |
| Test | Target/UOM | Range | Result |
| CONDUCTIVITY | < 1.0 UMHOS US/CM | | <1.0 UMHOS |
| Sample ID | | | |
| Tested | 02/09/24 | | |
| Tester | TBUR | | |
| Recorded | 02/09/24 | | |
| TOC < 50 PPB | < 50 PPB PPB | | <50 PPB |
| Sample ID | | | |
| Tested | 02/09/24 | | |
| Tester | TBUR | | |
| Recorded | 02/09/24 | | |
| SODIUM < 5 PPB | < 5 PPB PPB | | <5 PPB |
| Sample ID | | | |
| Tested | 02/09/24 | | |
| Tester | TBUR | | |
| Recorded | 02/09/24 | | |
| CHLORIDE < 5 PPB | < 5 PPB PPB | | <5 PPB |
| Sample ID | | | |
| Tested | 02/09/24 | | |
| Tester | TBUR | | |
| Recorded | 02/09/24 | | |
| SILICA < 3 PPB | 3 PPB OR LESS PPB | | 3 PPB OR LES |
| Sample ID | | | |
| Tested | 02/09/24 | | |
| Tester | TBUR | | |
| Recorded | 02/09/24 | | |
| HEAVY METALS AS LEAD | < 10 PPB PPM | | <10 PPB |
| Sample ID | | | |
| Tested | 02/09/24 | | |
| Tester | TBUR | | |
| Recorded | 02/09/24 | | |
| PPM NO3 | < 0.4 PPM PPM | | <0.4 PPM |
| Sample ID | | | |
| Tested | 02/09/24 | | |
| Tester | TBUR | | |
| Recorded | 02/09/24 | | |
| PPM SULFATE (SO4) | < 1.0 PPM PPM | | <1.0 PM |
| Sample ID | | | |
| Tested | 02/09/24 | | |
| Tester | TBUR | | |
| Recorded | 02/09/24 | | |



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

| | | |
|---------------------------------------|---------------------------|-------------|
| RESISTIVITY IN MEGOHM*CM @25 C | > 1.0 MEGOHM MEGOHM/CM | >1.0 MEGOHM |
| Sample ID | | |
| Tested | 02/09/24 | |
| Tester | TBUR | |
| Recorded | 02/09/24 | |
| CONDUCTIVITY (US/CM) | < 1 UMHOS US/CM | <1 UMHOS |
| Sample ID | | |
| Tested | 02/09/24 | |
| Tester | TBUR | |
| Recorded | 02/09/24 | |
| EXPIRATION DATE | MM/DD/YY MM/DD/YYYY | 02/28/2025 |
| Sample ID | | |
| Tested | 02/09/24 | |
| Tester | TBUR | |
| Recorded | 02/09/24 | |

Lot # 402902
Made 01/31/24

Chemical Product: 9766
 WATER DISTILLED ASTM TYPE II

| Test | Target/UOM | Range | Result |
|------------------------|---------------------|-------|------------|
| CONDUCTIVITY | <1.0 UMHOS US/CM | | <1.0 UMHOS |
| Sample ID | | | |
| Tested | 02/06/24 | | |
| Tester | TBUR | | |
| Recorded | 02/06/24 | | |
| TOC < 50 PPB | <50 PPB PPB | | <50 PPB |
| Sample ID | | | |
| Tested | 02/06/24 | | |
| Tester | TBUR | | |
| Recorded | 02/06/24 | | |
| SODIUM CONTENT | <5 PPB % | | <5 PPB |
| Sample ID | | | |
| Tested | 02/06/24 | | |
| Tester | TBUR | | |
| Recorded | 02/06/24 | | |
| CHLORIDE | <5 PPB CHLORID | | <5 PPB |
| Sample ID | | | |
| Tested | 02/06/24 | | |
| Tester | TBUR | | |
| Recorded | 02/06/24 | | |



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

| | | |
|---------------------------------------|------------------------------|-----------------|
| SILICA UG/L | <3 PPB UG/L (PPB) | <3 PPB |
| Sample ID | | |
| Tested | 02/06/24 | |
| Tester | TBUR | |
| Recorded | 02/06/24 | |
| HEAVY METALS (AS LEAD) | <10 PPB PPM | <10 PPB |
| Sample ID | | |
| Tested | 02/06/24 | |
| Tester | TBUR | |
| Recorded | 02/06/24 | |
| NITRATE (NO3) CONTENT | <0.4 PPM PPM | <0.4 PPM |
| Sample ID | | |
| Tested | 02/06/24 | |
| Tester | TBUR | |
| Recorded | 02/06/24 | |
| PHOSPHATE CONTENT | <1.0 PPM PPM | <1.0 PPM |
| Sample ID | | |
| Tested | 02/06/24 | |
| Tester | TBUR | |
| Recorded | 02/06/24 | |
| SULFATE (SO4) CONTENT | <1.0 PPM MAX 60 MG/KG | <1.0 PPM |
| Sample ID | | |
| Tested | 02/06/24 | |
| Tester | TBUR | |
| Recorded | 02/06/24 | |
| RESISTIVITY IN MEGOHM*CM @25 C | >1.0 MEG-OHM-CM MEGOHM/CM | >1.0 MEG-OHM-CM |
| Sample ID | | |
| Tested | 02/06/24 | |
| Tester | TBUR | |
| Recorded | 02/06/24 | |
| EXPIRATION DATE | Y MM/DD/YYYY | Y/N Y |
| Sample ID | | |
| Tested | 02/06/24 | |
| Tester | TBUR | |
| Recorded | 02/06/24 | |
| EXPIRATION DATE | MM/DD/YY MM/DD/YYYY | 02/28/2025 |
| Sample ID | | |
| Tested | 02/06/24 | |
| Tester | TBUR | |
| Recorded | 02/06/24 | |

Made 01/29/24



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 4

Date: 02/09/24 at 2:54 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager