



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 4132			
HYDROCHLORIC ACID 0.2 NORMAL IN IPA			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 02/09/24			
Tester ENOR			
Recorded 02/09/24			
<b>SAMPLE COLLECTION &amp; HANDLING</b>	YES		YES
<b>PER LAB040</b>	YES		
Sample ID			
Tested 02/09/24			
Tester ENOR			
Recorded 02/09/24			
<b>DATE OF SAMPLING FOR TEST</b>	MM/DD/YYYY		02/09/2024
<b>PURPOSES</b>	MM/DD/YYYY		
Sample ID			
Tested 02/09/24			
Tester ENOR			
Recorded 02/09/24			
<b>AMBIENT TEMPERATURE DURING</b>	25.0	10.0 - 35.0	20.3
<b>SAMPLING DEGC</b>	DEG C		
Sample ID			
Tested 02/09/24			
Tester ENOR			
Recorded 02/09/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		02/09/2024
	MM/DD/YYYY		
Sample ID			
Tested 02/09/24			
Tester ENOR			
Recorded 02/09/24			
<b>AMBIENT TEST TEMPERATURE</b>	20.0	15.0 - 30.0	20.0
<b>DEG C</b>	DEG C		
Sample ID			
Tested 02/09/24			
Tester ENOR			
Recorded 02/09/24			
<b>N.I.S.T. Traceable to SRM 84 (KHP)</b>	YES		YES
	YES		
Sample ID			
Tested 02/09/24			
Tester ENOR			
Recorded 02/09/24			
<b>Actual Normality</b>	0.2000	0.1991 - 0.2009	0.2005
	EQ/L		
Sample ID			
Tested 02/09/24			
Tester ENOR			
Recorded 02/09/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE APPEARANCE	PASS
Sample ID		
Tested	02/09/24	
Tester	ENOR	
Recorded	02/09/24	
<b>INSTRUMENTS USED DURING PREPARATION</b>	<b>LAB EQUIPMENT USED INSTRUMENT</b>	M115,VF3-J18,BID-06969,PID-89
Sample ID		
Tested	02/09/24	
Tester	ENOR	
Recorded	02/09/24	
<b>Actual Density</b>	0.79591 GM/ML	0.79476 - 0.79706 0.79622
Sample ID		
Tested	02/09/24	
Tester	ENOR	
Recorded	02/09/24	
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY	08/30/24
Sample ID		
Tested	02/09/24	
Tester	ENOR	
Recorded	02/09/24	
<b>Lot #</b>	4020815	
<b>Made</b>	02/09/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

  
**Quality Manager**