

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: THE255F SULFURIC ACID 0.002 NORMAL (N/500)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 02/08/24			
Tester MHUS			
Recorded 02/08/24			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 02/08/24			
Tester MHUS			
Recorded 02/08/24			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		02/08/2024
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 02/08/24			
Tester MHUS			
Recorded 02/08/24			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	22.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 02/08/24			
Tester MHUS			
Recorded 02/08/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		02/08/2024
	MM/DD/YYYY		
Sample ID			
Tested 02/08/24			
Tester MHUS			
Recorded 02/08/24			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.0
DEG C	DEG C		
Sample ID			
Tested 02/08/24			
Tester MHUS			
Recorded 02/08/24			
Actual Normality	0.0020	0.0018 - 0.0022	0.0020
	EQ/L		
Sample ID			
Tested 02/08/24			
Tester MHUS			
Recorded 02/08/24			
INSTRUMENTS USED DURING	LAB EQUIPMENT USED		VF3-J21,M55,BURET:3902,P17
PREPARATION	INSTRUMENT		2
Sample ID			
Tested 02/08/24			
Tester MHUS			
Recorded 02/08/24			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 02/08/24 at 3:52 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Density	0.99832	0.99732 - 0.99932	0.99829
	GM/ML		

Sample ID

Tested 02/08/24

Tester MHUS

Recorded 02/08/24

Lot # 4020768

Made 02/08/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager