

Page: 1

Date: 02/15/24 at 2:05 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	1795			
Ü		NE STANDARD 1000 PPM AS	CL2 (Refrigerate)	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	02/15/24			
Tester	MHUS			
Recorded	02/15/24			
PPM CHLORINE (CL2)		1000.0	995.0 - 1005.0	998.7
, ,		PPM		
Sample ID				
Tested	02/15/24			
Tester	MHUS			
Recorded	02/15/24			
APPEARANCE AND COL		CLEAR PALE YELLOW		PASS
		APPEARANCE		17.00
Sample ID		ALL ENLANCE		
Sample ID Tested	02/15/24			
	02/15/24 MHUS			
Tester				
Recorded	02/15/24	INICTOLIMENT		M110 V/F5 L15 DID 2002 DID 1
INSTRUMENTS USED D	URING	INSTRUMENT		M112,VF5-L15,BID-3902,PID-1
PREPARATION				77
Sample ID				
Tested	02/15/24			
Tester	MHUS			
Recorded	02/15/24			
Actual Density		1.00111	0.99936 - 1.00286	1.00086
		GM/ML		
Sample ID				
Tested	02/15/24			
Tester	MHUS			
Recorded	02/15/24			
EXPIRATION DATE		MM/DD/YY		04/20/24
		MM/DD/YYYY		
Sample ID				
Tested	02/15/24			
Tester	MHUS			
Recorded	02/15/24			
SHELF LIFE OF PRODU	СТ	2 MONTHS		2 MONTHS
		TIME		
Sample ID				
Tested	02/15/24			
Tester	MHUS			
Recorded	02/15/24			
Lot # 4021420				
Made 02/15	/ 44			



Page: 2

Date: 02/15/24 at 2:05 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager