Page: 1

Date: 02/12/24 at 5:06 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged		TA STANDARD 1 PPM AS NH3		
Test	74110101	Target/UOM	Range	Result
	O FORMULATION ON	YES		YES
FILE		YES		
S	Sample ID			
	Гested 02/12/24			
٦	Tester AROS			
F	Recorded 02/12/24			
CERTIFIED P	PM NH3	1.0	0.8 - 1.2	1.1
		PPM		
5	Sample ID			
	Гested 02/12/24			
٦	Tester AROS			
F	Recorded 02/12/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
9	Sample ID			
٦	Tested 02/12/24			
٦	Tester AROS			
F	Recorded 02/12/24			
INSTRUMENT	TS USED DURING	INSTRUMENT		P69,GC,T225
PREPARATIO	N			
9	Sample ID			
٦	Tested 02/12/24			
7	Tester AROS			
F	Recorded 02/12/24			
EXPIRATION DATE		MM/DD/YYYY		04/17/2024
		MM/DD/YYYY		
5	Sample ID			
٦	Tested 02/12/24			
٦	Tester AROS			
F	Recorded 02/12/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Pareland St