

Page: 1

Date: 02/12/24 at 2:37 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: 1795			
		NE STANDARD 1000 PPM AS	L2 (Refrigerate)	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	02/12/24			
Tester	AFRE			
Recorded	02/12/24			
PPM CHLORINE (CL2)		1000.0	995.0 - 1005.0	997.4
		PPM		
Sample ID				
Tested	02/12/24			
Tester	AFRE			
Recorded	02/12/24			
APPEARANCE AND CO		CLEAR PALE YELLOW		PASS
	-	APPEARANCE		. , .55
Sample ID				
Tested	02/12/24			
Tester	AFRE			
Recorded	02/12/24			
NSTRUMENTS USED I		INSTRUMENT		M103,VF5-L17
	DOMING	INO INCINICINI		W103, VF3-L17
PREPARATION				
Sample ID				
Tested	02/12/24			
Tester	AFRE			
Recorded	02/12/24			
Actual Density		1.00111	0.99936 - 1.00286	1.00044
		GM/ML		
Sample ID				
Tested	02/12/24			
Tester	AFRE			
Recorded	02/12/24			
EXPIRATION DATE		MM/DD/YY		04/17/24
		MM/DD/YYYY		
Sample ID				
Tested	02/12/24			
Tester	AFRE			
Recorded	02/12/24			
SHELF LIFE OF PRODI		2 MONTHS		2 MONTHS
		TIME		
Sample ID				
Tested	02/12/24			
Tester	AFRE			
Recorded	02/12/24			
Lot # 4020				
Made 02/12	2/24			

Page: 2

Date: 02/12/24 at 2:37 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager