

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

	<u> </u>			
Packaged Product:	6926			
	CONDUCT	FIVITY STANDARD 3,000 uMH		
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
	02/16/24			
Tester	JRIO			
	02/16/24			
CONDUCTIVITY @25 C.		3000 US/CM	2999 - 3001	3000
Sample ID				
	02/16/24			
	JRIO			
	02/16/24			
QCSP30:CONDUCTIVITY	,	YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
Tested	02/16/24			
Tester	JRIO			
	02/16/24			
TRACEABLE TO N.I.S.T.	SRM 3191	YES		YES
		YES		
Sample ID				
Tested	02/16/24			
Tester	JRIO			
	02/16/24			
APPEARANCE AND COL	.OR	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	02/16/24			
Tester	JRIO			
Recorded	02/16/24			
INSTRUMENTS USED DU	JRING	INSTRUMENT		M92,T229
PREPARATION				
Sample ID				
Tested	02/16/24			
Tester	JRIO			
Recorded	02/16/24			
EXPIRATION DATE		MM/DD/YY		02/28/25
		MM/DD/YYYY		
Sample ID				
Tested	02/16/24			
Tester	JRIO			
	02/16/24			
SHELF LIFE OF PRODUC	СТ	SHELF LIFE		1 YEAR
		TIME		
Sample ID				
	02/16/24			
Tester	JRIO			
Recorded	02/16/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4021648
Made	02/16/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

thand St

Quality Manager