02/20/24

Made

Page: 1

Date: 02/21/24 at 3:21 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Package	ed Product:	8710			
Ū			PHENATE REAGENT APHA&EPA	350.1 FOR AMMONIA	
Test			Target/UOM	Range	Result
PREPARE	D TO FORMU	LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	02/20/24			
	Tester	MHAY			
	Recorded	02/20/24			
APPEARANCE AND COLOR		LOR	CLEAR WATER WHITE		PASS
			LIQUID		
			APPEARANCE		
	Sample ID				
	Tested	02/20/24			
	Tester	MHAY			
	Recorded	02/20/24			
INSTRUMENTS USED DURING		URING	INSTRUMENT		M100,M108,VF500-X43
REPARA	ATION				
	Sample ID				
	Tested	02/20/24			
	Tester	MHAY			
	Recorded	02/20/24			
XPIRATI	ON DATE		MM/DD/YYYY		08/30/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	02/21/24			
	Tester	MHAY			
	Recorded	02/21/24			
10	ot# 40220	31			
_`					

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager