

Page: 1

Date: 02/19/24 at 11:43 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

## **Certificate of Analysis / QC Results**

Customer PO #:

Package	d Product:	MCH001F			
		Alkalini	ty Standard 100 ppm	(Special)	
Test			Target/UOM	Range	Result
	D TO FORMULA	TION ON	YES		YES
FILE			YES		
	Sample ID				
		2/19/24			
		HUS			
	Recorded 02	2/19/24			
Alkalinity			100.00	90.00 - 110.00	95.28
			ALKALINITY		
	Sample ID				
		2/19/24			
		HUS			
		2/19/24			
Actual Density		0.99833	0.99820 - 0.99846	0.99840	
			GM/ML		
	Sample ID				
		2/19/24			
	Tester M	HUS			
		2/19/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS	
			APPEARANCE		
	Sample ID				
	Tested 02	2/19/24			
	Tester M	HUS			
	Recorded 02	2/19/24			
INSTRUMENTS USED DURING		INSTRUMENT		M81,VF2-A13,M55,P168	
PREPARA <sup>*</sup>	TION				
	Sample ID				
		2/19/24			
	Tester M	HUS			
	Recorded 02	2/19/24			
SHELF LIFE OF PRODUCT		3 MONTHS		3 MONTHS	
			TIME		
	Sample ID				
	•	2/19/24			
		HUS			
		2/19/24			
EXPIRATION			MM/DD/YYYY		05/30/2024
			MM/DD/YYYY		
	Sample ID				
		2/19/24			
		HUS			
		2/19/24			
La	t# 4021508				
Mad	ue 02/19/2	. ¬			



Page: 2

Date: 02/19/24 at 11:43 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**