02/23/24

Made

Page: 1

Date: 02/23/24 at 1:56 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

_		TAN Wo	orking STANDARD 0.00435N ((0.26 ppm) IN KEROSENE	
Test			Target/UOM	Range	Result
PREPAREI	D TO FORMUL	ATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	02/23/24			
	Tester	GWEB			
	Recorded	02/23/24			
APPEARAI	NCE AND COL	_OR	CLEAR TO YELLOW LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	02/23/24			
	Tester	GWEB			
	Recorded	02/23/24			
INSTRUMENTS USED DURING			INSTRUMENT		P172,VF2-B17
PREPARA	TION				
	Sample ID				
	Tested	02/23/24			
	Tester	GWEB			
	Recorded	02/23/24			
EXPIRATION DATE			MM/DD/YYYY		08/30/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	02/23/24			
	Tester	GWEB			
	Recorded	02/23/24			
Lo	t# 40223	20			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager

hichard Stol