

Page: 1

Date: 02/23/24 at 9:20 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	3867			
	TURBIC	OITY STD. (FORMAZIN) 1.0 N	TU (APHA)	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	02/23/24			
Tester	AROS			
Recorded	02/23/24			
TURBIDITY (NTU)		1.00	0.95 - 1.05	1.00
		NTU		
Sample ID				
Tested	02/23/24			
Tester	AROS			
Recorded	02/23/24			
APPEARANCE AND CO		SLIGHTLY TURBID LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	02/23/24			
Tester	AROS			
Recorded	02/23/24			
INSTRUMENTS USED		INSTRUMENT		P45,VF1-Z36,P78,VF500-X42
PREPARATION				
Sample ID				
Tested	02/23/24			
Tester	AROS			
Recorded	02/23/24			
SHELF LIFE OF PROD		2 WEEKS		2 WEEKS
OHEER EN E OF TROOP	001	TIME		2 WEEKS
Sample ID		1 1141		
Tested	02/23/24			
Tester	02/23/24 AROS			
Recorded	02/23/24			
EXPIRATION DATE	02/23/24	MM/DD/YYYY		03/08/2024
EXPIRATION DATE		MM/DD/YYYY		03/00/2024
Sample ID				
Tested	02/23/24			
Tester	AROS			
Recorded	02/23/24			
Lot # 4022	348			
Made 02/2	3/24			



Page: 2

Date: 02/23/24 at 9:20 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager